

10/530602

JCO6 Rec'd PTO 07 APR 2005

**APPLICATION FOR
UNITED STATES PATENT**

by

Peter H. Breen, MD, FRCPC

assignor to

The Regents of The University of California

for

**BYMIXER APPARATUS AND METHOD FOR FAST-
RESPONSE, ADJUSTABLE MEASUREMENT
OF MIXED GAS FRACTIONS IN
VENTILATION CIRCUITS**

Prepared by Robert D. Buyan
STOUT, UXA, BUYAN & MULLINS, LLP
4 Venture, Suite 300
Irvine, CA 92618
949/450-1750 x220
fax: 949/450-1764

DOCKET NO. UCIVN-024US

Express Mail No. EV330332866US

**BYMIXER APPARATUS AND METHOD FOR FAST-RESPONSE,
ADJUSTABLE MEASUREMENT OF MIXED GAS
FRACTIONS IN VENTILATION CIRCUITS**

RELATED APPLICATION

This patent application claims priority to United States Provisional Patent Application Serial No. 60/417,982 entitled "Bymixer Apparatus for Fast-Response, Adjustable measurement of Mixed Expired Gas Fractions in the Anesthesia Circle Circuit and Related Method" filed on October 11, 1002, the entirety of which is
5 expressly incorporated herein by reference.

**STATEMENT REGARDING FEDERALLY SPONSORED
RESEARCH OR DEVELOPMENT**

Work connected with this invention was supported in part by the National Heart Lung and Blood Institute Grant R01 HL-42637. The United States Government
10 may have rights in this invention.

FIELD OF THE INVENTION

The present invention relates generally to biomedical devices and methods, and more particularly to devices and methods for anesthesia, critical care medicine, ventilation and monitoring of pulmonary function.

15 **BACKGROUND OF THE INVENTION**

Carbon dioxide (CO₂) is normally produced in the tissues of the human body where it becomes dissolved in the blood. The CO₂ is then transported in blood to the lung where it diffuses across alveolar membranes and is expelled from the lungs during exhalation.

20 The term "capnography" refers generally to the measurement of CO₂ in airway gas during the ventilation cycle. In patients who are undergoing anesthesia or mechanical ventilation, capnography is sometimes used to measure the partial

pressure of CO_2 (PCO_2) at the airway opening during the ventilation cycle. During the inspiratory phase of the ventilation cycle (i.e., inhalation), a flow of inspired respiratory gas passes through the airway opening. Such inspired respiratory gas typically contains little or no CO_2 . Thus, during the inspiratory phase, the capnograph
5 obtains an inspiratory baseline PCO_2 measurement of zero. During the second phase of the ventilation cycle (expiratory upstroke), alveolar gas from the respiratory bronchioles and alveoli begins to pass out of the patient's airway and the capnogram measures a rapid increase in CO_2 as the expiratory phase of the ventilation cycle proceeds. The third phase of the ventilation cycle is known as the "alveolar plateau,"
10 during which a relatively constant PCO_2 is measured at the airway opening. The PCO_2 of the expired respiratory gas at the end of this third phase of the ventilation cycle (PETCO_2) is typically of particular interest as it represents the last alveolar gas sampled at the airway opening during expiration. Finally, the fourth phase of the ventilation cycle is the inspiratory downstroke, during which the next inspiratory
15 phase begins.

While these direct capnographic measurements at the airway opening do provide the clinician with important diagnostic information, the usefulness of such information is limited due to the fact that direct capnographic measurements of this type merely measure the partial pressure of CO_2 without relating such measurement to
20 the volume of respiratory gas that is passing through the airway opening as the measurement is taken. In view of this shortcoming of traditional capnography, it is now believed that a measurement of volume-normalized average alveolar PCO_2 and pulmonary carbon dioxide elimination (\dot{V}_{CO_2}) are more clinically useful than the traditionally used end-tidal PCO_2 (PETCO_2).

Additionally, anesthesiologists, pulmonologists and critical care physicians are now beginning to consider another measurable variable known as “pulmonary carbon dioxide elimination per breath ($\dot{V}_{\text{CO}_2, \text{br}}$).” $\dot{V}_{\text{CO}_2, \text{br}}$ is arrived at by multiplication and integration of the airway flow and PCO_2 of the respiratory gas over all four phases of
5 the respiratory cycle.

Also, there is growing acceptance of a technique known as indirect calorimetry (e.g., the measurement and/or computation of CO_2 elimination and O_2 uptake) during anesthesia or mechanical ventilation for the rapid detection of various untoward states such as metabolic upset (e.g. onset of anaerobic metabolism) or
10 pulmonary embolism.

The measurement of pulmonary carbon dioxide elimination (\dot{V}_{CO_2}), pulmonary oxygen uptake (\dot{V}_{O_2}) and other indirect calorimetric measurements are facilitated by sampling of mixed respiratory gas. Such sampling of mixed respiratory gas may be accomplished in several ways. One way is to attach a collection vessel such as a bag
15 to the ventilation circuit to collect expired respiratory gas over a period of time. This collection technique is time consuming and of limited value because the collected mixture of respiratory gas is obtained from only one location in the ventilation circuit (e.g., from the expiratory flow conduit). Another technique for sampling mixed respiratory gas is through use of an in-line bymixer device. The bymixer devices of
20 the prior art have been constructed to continually divert a portion of respiratory gas flowing through a conduit into a reservoir. Sanjo, Y., Ikeda, K., *A Small Bypass Mixing Chamber for Monitoring Metabolic Rate and Anesthetic Uptake*, J. Clin. Monit. 1987; 3: 235-243; Breen P.H., Serina E.R., *Bymixer Provides On-Line Calibration of Measurement of Volume Exhaled Per Breath*, Ann. Biomed. Eng.

1997; 25:164-171. However, such prior art bymixers were typically difficult to construct and thus somewhat expensive. Also, the gas collection reservoirs of such prior art bymixers were of constant volume and the gas diverting tubes were of constant dimensions and, thus, could not be rapidly adapted or adjusted to
5 accommodate patients of varying size (e.g., small pediatric patients and large adult patients) or changes that may occur in a particular patient's ventilation parameters or clinical status. Finally, the gas collection reservoirs of the prior art bymixers were prone to collect condensed water vapor and respiratory debris and were difficult to clean.

10 Accordingly, there remains a need in the art for the development of a new bymixer device that is simple and economical to use and is adjustable or adaptable so as to be useable in patients of varying size (e.g., small pediatric patients and large adult patients) and to optimize the continuing measurements made during a given procedure as changes occur in the ventilation circuit and/or in a patient's ventilation
15 parameters or clinical status.

SUMMARY OF THE INVENTION

The present invention provides a new bymixer device and method for obtaining fast-response, accurate measurements of mixed expired gas fractions in various types of ventilation circuits, including open (non-rebreathing) circuits, closed
20 (rebreathing) circuits, circle (rebreathing with optional added fresh gas) circuits, etc. The bymixer is of a novel parallel design, which facilitates adjustable response, easy cleaning, and construction from standard airway circuit components. This bymixer may serve as a platform or enabling technology to facilitate further use of indirect calorimetry during anesthesia and/or mechanical ventilation in critical care settings.

In accordance with the present invention, there is provided a bymixer device that is connectable to a respiratory gas flow conduit in a ventilation circuit (e.g., an open circuit, a closed circuit, a circle circuit, etc.) that is used for ventilating a human or veterinary patient. In general, the bymixer device of the present invention
5 comprises a) a flow dividing manifold (e.g., a Y or T) for dividing the flow of respiratory gas into first and second flow streams, b) a main or direct flow channel that is connectable to the flow dividing manifold such that the first flow stream flows through the main flow channel, and c) a bypass flow channel that is connectable to the flow dividing manifold such that the second flow stream flows through the bypass
10 flow channel. The bypass flow channel includes a flow-restrictor (e.g., an orifice or other flow-restricting structure) that partially blocks the flow of respiratory gas through the bypass flow channel, a mixing chamber positioned upstream of the flow restrictor and a sampling apparatus (e.g., a port for withdrawing samples of gas from said mixing chamber and/or sensor(s) positioned within the mixing chamber).

15 Further in accordance with the present invention, all or a portion of the bymixer device (e.g., the mixing chamber) may be automatically or manually adjustable or variable in size such that the volume of respiratory gas contained in the mixing chamber may be varied. In combination with, or separately from, such adjustability in the size of the mixing chamber, the degree of flow restriction caused
20 by the flow restrictor may also be automatically or manually variable. The adjustability in mixing chamber size and/or degree of flow restriction allows the bymixer device to be adjusted or adapted in a manner that optimizes the rate of response in measurements made on mixed gas samples from the mixing chamber versus the homogeneity of the mixed gas samples obtained from the mixing chamber.
25 Also, such adjustability of mixing chamber volume and/or flow rate through the

mixing chamber allows the operator to adjust the bymixer to accommodate patients of varying size and/or to maintain optimal mixing of respiratory gas and monitoring of variables even when changes occur in the ventilation circuit or in a given patient's ventilation parameters and/or clinical status.

5 Still further in accordance with the present invention, one or more flow-disrupting surfaces and/or one or more mixing apparatus (e.g., mixing vanes, a rotating impeller, vibrating surface, moving member, etc.) may optionally be positioned in the mixing chamber to further enhance the mixing of respiratory gasses within the mixing chamber.

10 Further aspects of the present invention will become apparent to those of skill in the art upon reading and understanding the following detailed description and examples.

BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 is a schematic showing of a bymixer device of the prior art.

15 Figure 2A is a schematic diagram of a circle-type ventilation circuit incorporating a bymixer device of the present invention.

Figure 2B is a schematic diagram of an open ventilation circuit incorporating a bymixer device of the present invention.

20 Figure 2C is a longitudinal sectional view of a bymixer device of the present invention.

Figures 3A, 3B and 3C are graphs showing the correlation of bymixer mixed expired P_{CO_2} ($P\bar{E}_{CO_2}$) versus $P\bar{E}_{CO_2}$ measured in a mixed collection of expired gas, for mixing chamber volumes of 100, 150, and 200 ml. Each plotted point represents a steady state ventilation sequence of a CO_2 -producing lung simulator, over a range of

tidal volume (300-1200 ml) and respiratory frequency (6-20 breath/min); m, slope; b, Y-intercept; and R^2 , coefficient of determination (linear regression).

Figures 4A and 4B are graphs depicting data generated by Bland-Altman analysis. Figure 4A shows the difference between the bymixer P_{CO_2} and the value measured in the mixed collection of expired gas (exhaust gas collection bag) plotted against the average of the two values. Figure 4B shows the ratio of the bymixer P_{CO_2} -to-gas collection P_{CO_2} plotted against the average of the two values. (Dotted lines denote the $\text{mean} \pm 1.96$ standard deviations (SD), which encompass 95% of the measurement sequences. Each plotted point represents a steady state ventilation sequence of a CO_2 -producing lung simulator, over a range of tidal volume (300-1200 ml) and respiratory frequency (6-20 breath/min). Measurements for mixing chamber volumes of 100, 150, and 200 ml are combined.)

Figures 5A and 5B are graphs showing the effect of mixing chamber volume on bymixer P_{CO_2} during continuous sampling from the bymixer port (200 ml/min) by the side-stream gas analyzer. Data were digitally acquired at 100 Hz. Airway opening flow was processed by moving average filter over 7 data points to remove signal noise. For clarity, every 20th data point was plotted for bymixer P_{CO_2} . Relative to the flow signal, P_{CO_2} was advanced in time by transport delay, the time to aspirate gas through the sampling line. Transport delay was measured previously in a bench set-up. Respiratory frequency was 12 breath/min and tidal volume was 600 ml. With the mixing chamber volume of 100 ml (upper panel), oscillations in bymixer P_{CO_2} were about 1.3 mm Hg. The larger mixing chamber of 200 ml (lower panel) generated only tiny oscillations in bymixer P_{CO_2} of about 0.2 mm Hg.

DETAILED DESCRIPTION

The following detailed description, and the accompanying drawings to which it refers, are provided for the purpose of describing and illustrating certain examples

or specific embodiments of the invention only and not for the purpose of exhaustively describing all possible embodiments and examples of the invention. Thus, this detailed description does not in any way limit the scope of the inventions claimed in this patent application or in any patent(s) issuing from this or any related application.

5 As shown in Figure 1, a typical bymixer device of the prior art comprised a main flow conduit MFC that extends through a sealed mixing chamber MC. A right-angled upstream bypass tube UBT and a right-angled downstream bypass tube DBT extend through openings formed at longitudinally spaced-apart locations in the wall of the main flow conduit MFC, as shown. A porous baffle B divides the mixing
10 chamber MC into an upstream portion UP and a downstream portion DP. A fraction of the respiratory gas flowing through the main flow conduit MFC would enter the upstream bypass tube UBT and would flow therethrough and into the upstream portion UP of the mixing chamber MC. Samples of mixed respiratory gas could be withdrawn from the downstream portion DP of the mixing chamber MC into the
15 sample chamber SC where further mixing would occur and then through the gas sampling tube GST to the desired test apparatus where the desired analysis or measurement would be conducted. Mixed gas would also continually flow from the downstream portion DP of the mixing chamber MC, through the downstream bypass tube DBT and back into the main flow conduit MFC, thereby providing continual
20 turnover of respiratory gas within the mixing chamber MC. This prior art bymixer has several limitations. For example, the size of the mixing chamber was fixed. Thus, the volume of gas within the mixing chamber could not be varied to optimize mixing of the respiratory gas or to respond to variations in patient size, physiology or clinical status. Similarly, the size and dimensions of the bypass tubes BP were fixed
25 and could not be adjusted to vary the fraction of main flow into the mixing chamber

MC. Also, the diameter of the upstream bypass tube was smaller than the diameter of the main flow conduit MFC and its position within the main flow conduit MFC was fixed. Thus, the sampling of gas was always obtained from the same region (e.g., the center) of the main flow conduit and in situations where flow through the main flow conduit is laminar, certain fractions of such laminar flow (e.g., that flowing through the periphery of the conduit) could flow past the upstream bypass tube UBT and would not be included in the sample shunted into the mixing chamber MC.

In clinical practice, there is substantial variation in the body size, respiratory physiology and clinical status of patients. Even during the course of a single procedure (e.g., a surgical procedure wherein the patient is connected to an anesthesia/ventilation circuit) there may be variations in the patient's respiratory physiology and/or clinical status. However, because the prior art bymixer (Figure 1) had a mixing chamber MC of fixed size, it was not possible to adjust the size and/or volumetric capacity of the mixing chamber in response to such variations. Also, during manufacture, the positioning and securing of the upstream and downstream bypass tubes UBT, DBT was laborious and time consuming and the existence of these bypass tubes protruding into the lumen of the main flow conduit could, theoretically at least, result in trapping of condensed water vapor, microbes, mucoid matter or other contaminants.

The bymixer device 12 of the present invention, as shown in Figures 2A, 2B and 2C, overcomes some or all of the shortcomings of the prior art bymixer. As explained herebelow, this new bymixer 12 is relatively simple and inexpensive to manufacture, relatively devoid of in-line obstructions and may optionally be adjustable in ways that allow the mixing chamber volume and/or the rate of flow through the mixing chamber to be modified or adjusted, thereby accommodating

patients of varying body size, as well as differing or changing respiratory physiology and/or clinical status. Two types of ventilation circuits incorporating the bymixer 12 (a rebreathing circle circuit 10 and a non-rebreathing open circuit 10') are shown in Figures 2A and 2B, while details of the bymixer device 12 itself are shown in Figure 5 2C.

Specifically, Figure 2A shows an example of a circle ventilation circuit 10 which incorporates a bymixer 12 of the present invention. As shown, an airway device 14 such as an endotracheal tube, nasotracheal tube, tracheostomy tube, laryngeal mask airway or face mask is connected to the circle circuit 10 such that 10 respiratory gas will flow into and out of the patient's lungs L. The circle circuit 10 comprises an inspiratory flow conduit 30 having a one-way inhalation valve 32 and an expiratory flow conduit 34 having a one-way exhalation valve 36, as shown, to alternately allow inspiratory inflow and expiratory outflow into and out of the patient's lungs L. During inspiration, inspiratory respiratory gas flows through 15 inhalation valve 32, through inspiratory flow conduit 30, through the breathing device 14 and into the patient's lungs L. Thereafter, during expiration, respiratory gas is expelled from the patient's lungs L, through the airway device 14, through exhalation valve 36 and through the expiratory flow conduit 34. A pneumotachometer 29 and humidity/temperature sensor 28 are mounted near the airway device 14 to monitor 20 respiratory rate, gas flow, humidity and temperature. The humidity/temperature sensor 28 may be any suitable type of humidity sensor, such as that described in United States Patent No. 6,014,890 (Breen) entitled *Fast Response Humidity and Temperature Sensor Device*, the entirety of which is expressly incorporated herein by reference. In the example shown in Figure 2A, the bymixer 12 of the present 25 invention is attached to the expiratory flow conduit 34 such that expiratory respiratory

gas flowing through the conduit 34 will enter a flow dividing manifold 40 (e.g., a Y or T) which channels a portion of that flow into a main flow channel 42 and a portion of that flow into a bypass channel 44.

As may be seen clearly in the enlarged sectional view of Figure 2C, the bypass
5 channel 44 comprises a mixing chamber 46, a flow restrictor 50 (e.g., an orifice) positioned downstream of the mixing chamber 46 and a sampling port 48 for withdrawing samples of mixed gas from the mixing chamber 46. It will be appreciated, however, that sampling of gas within the mixing chamber 46 may alternatively be accomplished without the need for withdrawal of gas through a
10 sampling port by positioning one or more sensors (e.g., electrodes, optical sensors, chemical sensors, etc.) within the interior of the mixing chamber so that sample measurements may be made within the mixing chamber 46. Respiratory gas that has passed through the main flow channel 42 and bypass channel 46 then enters a flow combining manifold 54 (e.g., another Y or T) where it is recombined and continues
15 through the downstream portion of the expiratory flow conduit 34.

In some embodiments of the invention, a monitoring device 43 may be connected to the sampling port 48 by a sampling tube 41 such that continuous or periodic samples of mixed gas may be withdrawn from the sample port 48 into the monitoring device 43. The monitoring device 43 may be operative to analyze,
20 measure and/or otherwise determine any desired variables, such as volume-averaged expired PCO_2 ($P_{E_{CO_2}}$), pulmonary carbon dioxide elimination (\dot{V}_{CO_2}), pulmonary oxygen uptake (\dot{V}_{O_2}), concentration or partial pressure of volatile gases (e.g., anesthetic gases), concentration or partial pressure of non-volatile gases, and/or other variables. In some instances, the monitoring device may incorporate a computer,

microprocessor or other calculating apparatus and may be programmed or otherwise adapted to calculate one or more calculated values of interest based on one or more of the measured variables. In some instances, additional data such as patient body weight, barometric pressure, airway opening gas humidity and temperature, airway opening gas flow etc. may be input into the monitoring device 43, either by manual input (e.g., via a keyboard) or may be communicated by hard wired or wireless connection between the monitoring device 43 and one or more sensing apparatus operative to measure such values. These additional data (when obtained) may also be used in calculating some calculated values of interest. Examples of calculated values that may optionally be provided by the monitoring device include but are not limited to: pulmonary carbon dioxide elimination (\dot{V}_{CO_2}), pulmonary carbon dioxide elimination per breath ($V_{CO_2,br}$), pulmonary oxygen uptake (\dot{V}_{O_2}), pulmonary oxygen uptake per breath ($V_{O_2,br}$), minute ventilation (\dot{V}_E), tidal volume (V_T), Vital Capacity (VC), etc. The monitoring device 43 may include, or may be connected to, a display for displaying the measured variables and/or or computed values. For example, a waveform display 45 may provide a display of one or more waveforms such as flow, pressure, capnography, oxygen concentration, spirometry, etc. and/or one or more alphanumeric displays 47 (e.g., LED displays) may display numerical values and/or letters relating to variables or computed values such as such as those stated hereabove determined using the bymixer, as well as others such as peak inspiratory pressure, end tidal CO_2 , expired O_2 , inspired O_2 , CO_2 elimination, \dot{V}_{CO_2} , and O_2 uptake, \dot{V}_{O_2} , tidal volume, minute volume, airway pressure, airway compliance, estimated energy required (EER), respiratory quotient (RQ), etc. In particular, the Respiratory Quotient (RQ) is a determined parameter calculated from the quotient of pulmonary carbon

dioxide elimination (\dot{V}_{CO_2}) and pulmonary oxygen uptake (\dot{V}_{O_2}). The Respiratory Quotient is a very sensitive indicator of the metabolic state of the patient, particularly in its ability to detect the change from aerobic to anaerobic metabolism. The Respiratory Quotient depends only upon inspired and mixed expired gas fraction
5 measurements for its calculation, and in particular, does not require any flow measurements. Thus, with a bymixer 12 on the expiratory flow conduit 34 (e.g., expired air limb of the ventilation circuit) and possibly, as necessary, another bymixer 12 on the inspiratory flow conduit 30 (e.g., the inspiratory limb of the ventilation circuit), the Respiratory Quotient can be measured and determined without the need
10 for any flow measuring device.

One example of a monitoring device 43 that may be used is the Capnomac Ultima available commercially from Datex Medical Instruments, Instrumentarium Corp., Helsinki, Finland. Another example is the Datex-Ohmeda Division,
15 Instrumentarium Corp. (Helsinki, Finland; Madison, WI) Airway Module, M-CAiOVX (gas composition/indirect calorimetry) and S/5 portable compact monitor, which can measure airway oxygen uptake and carbon dioxide elimination. The bymixer apparatus 12 can provide accurate, on-line and simultaneous measurements of \dot{V}_{CO_2} and \dot{V}_{O_2} to provide calibration values for the M-CAiOVX
20 measurements. In at least some applications it may be of value to have a bymixer 12 of the present invention in a ventilation circuit in addition to another device for breath-by-breath calorimetric measurements at the patient's airway, such as the Datex-Ohmeda Airway Module, M-CAiOVX (gas composition/indirect calorimetry) and S/5 portable compact monitor. In this regard, the simultaneous side stream
25 analysis/measurement of gas fractions at the airway opening, along with a

measurement of flow at the airway opening (such as with a pneumotachometer) will allow the generation of breath-by-breath measurements of $\dot{V}O_2$, $\dot{V}O_{2,br}$ and $\dot{V}CO_2$, and will allow the determination of $\dot{V}CO_{2,br}$ and $\dot{V}O_{2,br}$ via the online multiplication and integration on these gas fraction and flow values measured at the airway opening.

5 The bymixer flow measurement of $\dot{V}O_2$ during steady state can be used to calibrate these more unstable breath-by-breath measurements of indirect calorimetry. Then, during non-steady state conditions, were the bymixer flow measurement may not react fast enough to changes in patient pathophysiology, the breath-by-breath measurements maybe used to follow the patient's condition.

10 In some embodiments, the system may be fully or partially automated. For example, the system may include a programmable controller (e.g., a microprocessor or computer) that may receive input signal(s) from the monitoring device 48 and, in response to predetermined changes in measured variables or calculated values, may issue control signals to certain components of the system that are equipped to undergo
15 changes in response to such automated control signals. For example, such automated control system may optimize the size of the mixing chamber 46 and/or the diameter of the flow restricting orifice 50. In some applications, it may suggest or facilitate changes in ventilator settings, such as FiO_2 , respiratory rate, tidal volume, and positive end expiratory pressure (PEEP), to remedy undesirable changes or trends in
20 measured variables or computed values.

Also, in some embodiments, the monitoring device 43 may also include one or more auditory or visual alarms that will be triggered when certain measured variables and/or computed values move outside of preset limits.

In the circle ventilation circuit 10 shown in Figure 2A, the portion of the

expiratory flow conduit downstream of the bymixer 12 flows into a CO₂ absorber 38 (e.g., SODASORB[®] 4-8 IND N MED, Daerx[®] Container Products, Cambridge, MA or ThermHOAbsorb[™], Raincoat Industries, Inc., Louisville, KY) which removes CO₂. After exiting the CO₂ absorber 38, the expiratory flow (less CO₂ absorbed by the absorber 38) may be mixed with fresh gas (e.g., air and/or oxygen and/or nitrogen and/or anesthetic gas(es)) entering through fresh gas inlet 18, and flows through the inspiratory flow conduit 30 and back into the patient's lungs L as described above. This circle (rebreathing) ventilation circuit 10 may be used in various settings including during anesthesia where it is desired to recycle volatile or gaseous anesthetics and in certain other types of mechanical ventilation (including anesthesia) where it is desirable to prevent loss of the temperature and humidity of the expired respiratory gas and/or where it is undesirable to allow the expired respiratory gas to escape into the surrounding room air.

The open ventilation circuit 10' shown schematically in Figure 2B includes many of the same components as the circle circuit 10 of Figure 2A. However, in this open circuit 10', the expired respiratory gas is allowed to vent out of the circuit 10', downstream of the bymixer and only fresh respiratory gas enters the inspiratory flow conduit 30. This open (non-rebreathing) circuit 10' is typically used in mechanical ventilation of non-anesthetized critical care patients or during surgical procedures in which volatile or gaseous anesthetics are not used. Although Figures 2A and 2B show ventilation circuits 10, 10' in which the bymixer 12 is located on the expiratory flow conduit 34, it will be appreciated that the bymixer 12 may also be located on the inspiratory flow conduit 30 to obtain time-averaged or mixed samples of inspiratory respiratory gases.

Optionally, in some embodiments of the invention, the mixing chamber 46 is

of variable size. This may be accomplished by constructing the mixing chamber of common corrugated tubing or other expandable or telescoping tubing such that the size and/or internal volume of the mixing chamber 46 may be varied. The advantages and clinical utility of this aspect of the invention are described more fully herebelow where reference is made to certain experimental data showing the utility of this feature as shown in Figures 3A-4B.

Optionally, in some embodiments of the invention, the flow restricting orifice 50 may be of variable size or diameter to permit the operator to easily adjust the flow rate of gas through the mixing chamber to optimize mixing.

10 Validation and Testing of Bymixer

Theoretical Background

Measurement of mixed expired gas concentrations is an essential component of the methodology to measure CO₂ elimination (\dot{V}_{CO_2}) and pulmonary oxygen uptake (\dot{V}_{O_2}) at the airway opening (1). In the normal condition where CO₂ is absent from inspired gas, \dot{V}_{CO_2} is given by

$$\dot{V}_{CO_2} = \dot{V}_E \cdot F\bar{E}_{CO_2} \quad (\text{Eq. 1})$$

where \dot{V}_E is the expired ventilation and $F\bar{E}_{CO_2}$ is the mixed expired CO₂ fraction.

On the other hand, \dot{V}_{O_2} is the difference between inspired and expired O₂ volumes, as given by

$$\dot{V}_{O_2} = \dot{V}_I \cdot F_{I_{O_2}} - \dot{V}_E \cdot F\bar{E}_{O_2} \quad (\text{Eq. 2})$$

where I denotes inspiration. Because expired volume is increased by increased temperature (T) and added water vapor (increased humidity), volumes must be corrected to standard temperature and pressure, dry (STPD) conditions or the error in \dot{V}_{O_2} can approach 50% as $F_{I_{O_2}}$ increases to unity. Because accurate differences between \dot{V}_I and \dot{V}_E are difficult to measure, the Haldane transformation is usually used, invoking conservation of the inert gas, nitrogen ($\dot{V}_I \cdot F_{I_{N_2}} = \dot{V}_E \cdot F\bar{E}_{N_2}$). By substitution into Eq. 2, \dot{V}_{O_2} can be expressed as a function of only \dot{V}_E , where:

$$\dot{V}_{O_2} = \dot{V}_E \cdot (F_{IO_2} \cdot F_{\bar{E}N_2}/F_{IN_2} - F_{\bar{E}O_2}). \quad (\text{Eq. 3})$$

Regardless of whether T and humidity differences between inspiration and expiration are managed by the Haldane transformation (Eq. 3) or by separate measurements of airway T and relative humidity (RH) (Eq. 2), the determination of \dot{V}_{CO_2} and \dot{V}_{O_2} requires measurements of mixed expired and inspired gas fractions. The classic method to obtain mixed expired gas fractions is to collect exhaled gas over a number of breaths in a collection chamber connected to the expiratory outlet of the ventilator. However, expired gas collection cannot be conducted in the anesthesia semi-open or closed anesthesia circle ventilating circuit because expired gas passes through a CO₂ absorber to become the next inspiration. Instead, to measure mixed expired gas fractions in the circle circuit, Applicant and other workers have used an inline bypass mixing chamber (e.g., a bymixer as shown in Figure 1). The term “bymixer” is named for the *by*-pass of a constant fraction of total flow through a *mix*-ing chamber. However, the response time of that bymixer is long and fixed, the mixing chamber is difficult to fabricate, clean and sterilize, and the device is bulky.

To solve these problems, the bymixer 12 of the present invention may be constructed from common anesthesia circuit components (Figure 2C). Instead of diverting gas flow into a separate, large mixing chamber, the new clinical bymixer incorporates a novel parallel tubing design. A constant fraction of total \dot{V} is diverted through the mixing chamber 46 (e.g., corrugated collapsible/expandable pediatric anesthesia circuit tubing), whose volume can be adjusted (e.g., by collapsing or expanding the corrugated tubing). The resistor 50 controls the fraction of bypass \dot{V} to total flow. As gas passes through the mixing chamber 46, it mixes longitudinally in the tubing. Flow-averaged mixed gas is sampled at the sampling port 48 for analysis by a side-stream sampling monitor 43.

In this study, the following questions were tested and answered: Is a constant fraction of main gas flow diverted through the mixing chamber (mandatory for mixed bypass gas samples to accurately represent total gas flow)? Does the longitudinal design of the tubular mixing chamber provide adequate mixing (no significant breath-

to-breath variation of mixed gas fractions)? What is the fastest but still accurate response of the new clinical bymixer when the mixing chamber volume is decreased (shortest length of mixing tubing)? Does continuous side-stream sampling flow rate affect the measurement of mixed gas fraction?

5 In order to test the performance of the new bymixer 12 during cyclical changes in gas fractions under actual expiratory flow conditions, the bymixer 12 was interposed in the expiratory flow conduit 34 of a ventilation circuit that was attached to a CO₂-producing metabolic lung simulator as described in United States Provisional Patent Application Serial No. 60/417,982, which is expressly incorporated
10 herein by reference and in Rosenbaum, A, and Breen, P.H., *Novel, Adjustable, Fast Response Bymixer Measures Mixed Expired Gas Concentrations In Circle Circuit*, Anesth Analg; 97:pp. (2003). Measurement of bymixer mixed expired P_{CO₂} was compared to the value in a gas collection from the exhaust port of the open circuit ventilator. The use of the metabolic lung simulator was mandatory for the execution
15 of this study, in order to provide a wide and controlled range of tidal volume, respiratory frequency (f), and mixed expired P_{CO₂}.

METHODS

Design and construction of the new clinical bymixer).

The bymixer 12 divides incoming total gas flow into two parallel channels, a main flow channel 42 and a bypass flow channel 44. In the bymixer 12 used in this experiment, the main flow channel 42 was constructed of a 24 cm length of standard 5 3/4 inch PVC pipe (22 mm ID). The mixing chamber 46 was a length of expandable/collapsible pediatric anesthesia circuit tubing (15 mm ID, Expandoflex, Cleveland Tubing Inc., Cleveland, TN). This adjustable tubing was connected, in series, to the sampling port 48 which was constructed from a sampling port adapter available commercially from Datex-Engstrom Division, Instrumentarium Corp., 10 Helsinki, Finland, a flow resistor 50, and a 12 cm length of standard 3/4 inch PVC tubing. The flow resistor 50 was constructed by drilling a 4 mm diameter hole in a plastic cap (NAS-820-10, Niagra Plastics, Erie, PA), placed inside a connector (Multi Adapter, Hudson RCI, Temecula, CA; 15 mm ID, 22 mm OD). In this study, the adjustable tubing lengths were 50, 65.5, and 121 cm, which generated mixing 15 chamber volumes (measured up to the sampling port) of 100, 150, and 200 ml, respectively. The volume of the bypass channel 44 from the sampling port 48 to the downstream "Y" connector 54 was 53 ml. The main flow channel 42 and bypass flow channel 44 were connected at each end by identical "Y" connectors 40, 54 (supplied 20 with standard anesthesia circle circuits). Volumes of channel components were determined by water displacement.

Determination of time constant (bymixer response).

The anesthesia monitor (Capnomac Ultima, Datex Medical Instruments, Instrumentarium Corp., Helsinki, Finland) sampling line was connected to the bymixer sampling port (200 ml/min) and the pneumotachometer adapter was attached 25 to the inlet of the bymixer. F_{O_2} (paramagnetic) and bymixer total flow were continuously captured (100 Hz) by analog-to-digital (A/D) acquisition PC card (DAQcard 700, National instruments, Austin, TX) installed in a notebook computer (Inspiron 3800, Dell Computer Corp., Austin, TX). The digital data acquisition

system was driven by a custom program (Delphi Pascal, Borland International, Scotts Valley, CA) written by our computer support specialist (David Chien) and one author (PHB). The bymixer was flushed with air to provide a baseline F_{O_2} of 21%. At time zero, oxygen flow of either 4, 8, or 12 L/min was abruptly connected to the bymixer input. The time constant (τ) was the time interval, from time zero, until F_{O_2} increased to 63% of its maximal value (8). The time constant was corrected for the F_{O_2} transport delay (2.95 sec) down the side-stream sampling system (9).

The bymixer response was also tested in the exhalation limb during mechanical ventilation of the CO_2 -producing metabolic lung simulator. Steady state ventilation was established at minute ventilation of 4, 8, or 12 L/min (respiratory frequency, f , was 10/min and inspiration-to-expiration time ratio, I:E, was 1:2). The bymixer 12 (100 ml mixing chamber) was separately flushed with air. During the inspiratory phase, the bymixer 12 was abruptly interposed in the expiratory flow conduit 34 of the ventilation circuit (time zero). Because bymixer data during ventilation was periodic and available only during expiration, Applicant used the time for bymixer P_{CO_2} (infra-red analysis) to reach 95% of its maximum value. This time for 95% response was corrected for the P_{CO_2} transport delay (1.76 sec) down the side-stream sampling system.

Validation of the accuracy of the new clinical bymixer.

To test the bymixer 12, Applicant used a modification of the metabolic lung simulator bench setup. The commercial lung simulator (Dual Adult TTL, Model 1600, Michigan Instruments, Inc., Grand Rapids, MI) generated a physiologic ventilation waveform by combining airway resistance elements to a bellows (residual volume=920 ml), whose compliance can be adjusted by springs. The mechanical lung was connected by a circular circuit to a metabolic chamber (airtight 18.6 L pail). Carbon dioxide was continuously infused (200 ml/min) by calibrated rotameter into the metabolic chamber. A fan and a baffle system inside the chamber ensured a homogeneous gas mixture. An occlusion roller pump (15 mm ID tubing; Precision Blood Pump, COBE Perfusion system, Lakewood, CO.) generated constant gas flow

(5 L/min) between the metabolic chamber and the mechanical lung. The mechanical lung was ventilated with 30% oxygen (Servo Ventilator 900C, Siemens, Sweden). The bymixer 12 was interpolated in the expiratory limb of the open circuit (no rebreathing).

5 For each length of mixing chamber expandable tubing, the bymixer 12 was tested during different ventilatory patterns, encompassing combinations of tidal volume (300-1200 ml) and respiratory f (6-20 breath/min). I:E ratio was 1:2. Gas was continuously sampled from the bymixer by the side-stream capnometer (bymixer $\text{P}\bar{\text{E}}_{\text{CO}_2}$). Before measurements began at each ventilator setting, steady state was confirmed
10 by stable values of P_{ETCO_2} and bymixer $\text{P}\bar{\text{E}}_{\text{CO}_2}$. A measurement sequence consisted of continuous digital acquisition of bymixer $\text{P}\bar{\text{E}}_{\text{CO}_2}$ and simultaneous collection of expired gas in a 15 L gas-impermeable collection bag (Hans Rudolph, Kansas City, MO) connected to the ventilator exhaust port. Measurements were conducted for 3 min (higher minute ventilation) to 5 min (lower minute ventilation). After the
15 measurement sequence, the expired gas collection was mixed by shaking and agitating small balls inside the bag. Gas collection $\text{P}\bar{\text{E}}_{\text{CO}_2}$ was measured by attaching the side-stream sampling line to a stop-cock on the collection bag. Prior to each measurement sequence, the collection bag was emptied by vacuum to prevent gas dilution error. After attaining steady state and just before gas collection began, the dead space of the
20 bag was flushed with exhaled gas from the ventilator exhaust port. Time-averaged bymixer $\text{P}\bar{\text{E}}_{\text{CO}_2}$ was compared to the value measured in the simultaneous expired gas collection.

Effect of tidal volume and respiratory frequency on oscillations of bymixer $P\bar{E}_{CO_2}$.

Using the above *Validation* experimental setup and measurement sequences in the bymixer (150 ml mixing chamber volume), Applicant conducted two additional protocols which measured oscillations of bymixer $F\bar{E}_{CO_2}$. First, respiratory frequency
5 was held constant (10 br/min) and tidal volume was varied from 300 to 1200 ml. Second, tidal volume was held constant (900 ml) and respiratory frequency was varied from 6 to 20 br/min.

Effect of intermittent (instead of continuous) sampling from the bymixer port.

Applicant conducted an additional validation protocol using the bymixer 12
10 with its mixing chamber 46 volume set at 150ml.). Gas was intermittently sampled from the bymixer sampling port 48, by manipulation of a 3-way stopcock, for short periods (about 3 sec). Several intermittent samples from the bymixer 12 were averaged for comparison with the expired gas collection (3-5 min), at each ventilator setting of tidal volume and frequency.

15 **Data Analysis.**

Bymixer bypass flow (\dot{V}_{BYPASS}) was calculated by

$$\dot{V}_{BYPASS} = V_{BYPASS} / \tau, \text{ Eq. 4}$$

where, V_{BYPASS} was the volume of the mixing chamber (measured up to the sampling port), and τ was the measured time constant of the bymixer (4,8). Then,

20
$$\text{Bymixer bypass ratio} = \dot{V}_{BYPASS} / \dot{V}_{TOTAL}, \text{ Eq. 5}$$

where \dot{V}_{TOTAL} was the total gas flow entering the bymixer.

In the validation of average bymixer P_{CO_2} versus the value measured in the expired gas collection (metabolic lung simulator),

$$\text{bymixer } P\bar{E}_{CO_2} = (\int_{t_0}^{t_{end}} P_{CO_2}(t) \cdot dt) / (t_{end} - t_0), \text{ Eq. 6}$$

25 where dt is the digital sampling interval (1/100 Hz) and t_0 and t_{end} were the beginning and end sampling times (sec) of bymixer P_{CO_2} .

Bymixer $P\bar{E}_{CO_2}$ were compared to expired gas collection $P\bar{E}_{CO_2}$ by least squares linear regression (slope, Y-intercept, and coefficient of determination, R^2) and by the limits of agreement technique described by Bland and Altman. Differences between

groups were sought by t-test or by analysis of variance (ANOVA). Computer programs were used for data analysis (Excel spreadsheet, Microsoft Corp., Redmond, WA), statistical testing (SigmaStat, SPSS, Chicago, IL), and graphical presentation (SigmaPlot 8.0, SPSS).

5

RESULTS

Figures 3A-C display the excellent linear regression correlation between bymixer $P\bar{E}_{CO_2}$ and the value measured in the expired gas collection over a wide range of $P\bar{E}_{CO_2}$ (6-50 mm Hg), for the bymixer with mixing chamber volumes set to 100, 150, and 200 ml. There was no significant difference in bymixer $P\bar{E}_{CO_2}$ accuracy among the mixing chamber volumes (ANOVA analysis of the $P\bar{E}_{CO_2}$ differences between the bymixer and expired gas collection measurements). For the bymixer set to mixing chamber volume of 150 ml, there was no significant difference in $P\bar{E}_{CO_2}$ accuracy between continuous and intermittent sampling from the bymixer port.

15 The Bland-Altman analysis derived the Limits of Agreement (LOA) as 0.07 ± 0.93 mm Hg (Figure 4A). Measurements for mixing chamber volumes of 100, 150, and 200 ml were combined. Inspection of the graph revealed that the P_{CO_2} difference, between the bymixer measurement and the simultaneous value measured in the expired gas collection, increased as the measurement increased along the x-axis. To correct for this effect, Figure 4B plotted the P_{CO_2} ratio (bymixer / bag) versus the average of the two values. The calculation of LOA (1.00 ± 0.03) demonstrated that 95% of the bymixer P_{CO_2} measurements were within 3% of the expired gas collection value.

25 Figures 5A and 5B displays the breath-by-breath oscillations in P_{CO_2} measured during continuous aspiration from the bymixer into the side-stream sampling gas analyzer. Oscillations in P_{CO_2} were larger with the smaller bymixer mixing chamber volume. Table 1, below, displays that the average P_{CO_2} oscillations increased from 0.1 to 0.7 mm Hg as bymixer mixing chamber volume decreased from 200 to 100 ml. The plot of P_{CO_2} oscillation (mm Hg) versus V_T (ml) (constant f) generated a significant

direct relationship (slope=0.0016; Y-intercept=-0.69; $R^2=0.92$). The plot of P_{CO_2} oscillation (mm Hg) versus f (min^{-1}) (constant V_T) resulted in a significant inverse relationship (slope=-0.062; Y-intercept=1.22; $R^2=0.91$). Thus, P_{CO_2} oscillations increased in magnitude as V_T increased and f decreased.

5 The ratio of bypass flow to total flow was similar (1:9) for the 3 mixing chamber volumes (Table 1). The time constant (τ) of the bymixer response to a change in input gas concentration (at 8 L/min) ranged from 6.4 to 14.1 sec for the smallest (100 ml) to largest (200 ml) mixing chamber volumes. Tripling of the time constant predicts 95% response. During minute ventilation of the metabolic lung
10 simulator at 4, 8, and 12 L/min, the times for 95% response of the bymixer (100 ml volume) were 19.0, 12.6, and 6.6 sec, respectively, significantly less than the values of 3τ (Table 1).

DISCUSSION

The bymixer 12 of the present invention incorporates a new design, compared
15 to the classic bymixer. Instead of diverting a portion of main flow into a surrounding reservoir in the classic bymixer (e.g. the prior art bymixer shown in Figure 1), the bymixer 12 of the present invention diverts a fraction of main flow through a parallel, longitudinal and adjustable mixing chamber 46. The flow resistor 50 (variable orifice) provided an easy control of fraction of bypass flow. To provide an accurate mixed
20 average gas fraction of total flow, the ratio of bypass flow/total flow must remain constant and gas must adequately mix by the time it reaches the sampling port 48. Figures 3A, 3B and 3C demonstrate the excellent correlation of bymixer $P_{\bar{E}CO_2}$ compared to the simultaneous value measured in the expired gas collection, over a wide range of V_T , f , and P_{CO_2} . The Bland-Altman Limits of Agreement analysis
25 shown in Figures 4A and 4B demonstrates excellent bymixer measurement accuracy, where 95% of the bymixer measurements were within 3% of the simultaneous value measured in the mixed expired gas collection. If present, the small bymixer P_{CO_2} oscillations (Figure 5 and Table 1) were time-averaged and did not degrade bymixer performance for mixing chamber volumes of 100, 150, and 200 ml.

For these mixing chamber volumes, the ratio of bypass flow to total flow was similar (1:9) because the major impedance to gas flow was the flow resistor. Increased length of the large bore tubing that formed the mixing chamber 46 tubing did not materially add to bypass flow resistance. Thus, dynamic response of the bymixer 12
5 can be improved by decreasing the volume of the mixing chamber 46 (e.g, in this example, by decreasing the length of the tubing). The data shown in Table 1 suggests that bymixer dynamic response (at 8 L/min) could be improved, beyond (less than) the 9.3 sec time constant of the bymixer with 100 ml mixing chamber, by further decreasing the volume of the mixing chamber 46. However, at some point, time-
10 averaging of increasing F_{CO_2} oscillations would significantly depart from the flow-averaged value and degrade bymixer accuracy. Interestingly, compared with constant gas flows (Table 1), the bymixer 12 demonstrated much faster response during mechanical ventilation of the metabolic lung simulator, presumably because the periodic, peak expiratory flows enhanced gas mixing in the bymixer 12.

15 There was no difference in bymixer accuracy between continuous and intermittent aspiration at the sampling port. Accordingly, the down-stream volume (measured from the sampling port) of the bypass channel was sufficiently large so that side-stream sampling (200 ml/min) did not spuriously sample gas from the main flow outlet during inspiration (when gas flow through the bymixer was zero).

20 The small bymixer F_{CO_2} oscillations, when present, represented slight incomplete mixing in bypass flow. F_{CO_2} oscillations decreased with smaller V_T because the ratio of V_T -to-mixing chamber volume decreased. F_{CO_2} oscillations decreased with higher f (at constant V_T) because increased overall gas flow (and velocity) improved gas mixing. The corrugations of the mixing chamber tubing
25 presumably added to gas mixing. The presence of bymixer P_{CO_2} oscillations was not significant, since simple time-averaging of the oscillations resulted in excellent bymixer accuracy (Figures 3A-C and 4A-B).

In summary, the novel, parallel design of the bymixer 12 provides accurate measurement of mixed expired gas fractions in the anesthesia circle circuit. Simple

changes in mixing chamber volume allow adjustable bymixer response time. The fast bymixer response (time constant = 6.4 sec) should permit measurements to be updated every 20 sec (where 95% response occurs by 3 time constants). The bymixer 12 of the present invention can be constructed from standard anesthesia circuit components, attaches easily to the anesthesia machine inspired outlet and expired inlet ports, is simple to clean and sterilize, and has no reservoir that can trap condensed water vapor from expired gas. This new bymixer 12 may facilitate more widespread use of indirect calorimetry (\dot{V}_{O_2} and \dot{V}_{CO_2}) during anesthesia and the non-invasive detection of metabolic upset (e.g. onset of anaerobic metabolism) and critical events (e.g. onset of pulmonary embolism).

Table 1. Effect of mixing chamber volume (V) on selected parameters of the new clinical bymixer (see Fig. 2).

Mixing Chamber V (ml)	τ (4 L/min) (sec)	τ (8 L/min) (sec)	τ (12 L/min) (sec)	Bypass Flow Ratio	F_{CO_2} Oscillations (mmHg)	slope	Y- intercept	R^2
100	14.1	6.4	4.0	1:8.6	0.68±0.89	0.995	0.04	0.9984
150	24.7	9.0	6.3	1:9.1	0.57±0.66	1.001	0.05	0.9997
200	32.3	14.1	9.1	1:9.7	0.14±0.28	1.010	-0.12	0.9983

τ was the time constant of the bymixer response to a change in input oxygen concentration, measured during three constant flow rates of O_2 . Bypass flow ratio = bypass flow/total flow. F_{CO_2} oscillations were measured during continuous aspiration from the bymixer sampling port. Slope, Y-intercept, and R^2 (coefficient of determination) characterized the correlation of bymixer mixed expired F_{CO_2} ($F_{E_{CO_2}}$) and the value measured in a simultaneous collection of expired gas (Figures 3A-3C), during ventilation of the CO_2 -producing lung simulator.

The foregoing detailed description and examples of various embodiments set forth hereabove shall not be construed in a limiting sense. The detailed description and examples are provided for purposes of illustration and description. It should be understood that all aspects of the invention are not limited to the specific depictions,

configurations or relative proportions set forth herein which depend upon a variety of conditions and variables. The specification is not intended to be exhaustive or to limit the invention to the precise forms disclosed herein. Various modifications and insubstantial changes in form and detail of the particular embodiments of the disclosed invention, as well as other variations of the invention, will be apparent to a person skilled in the art upon reference to the present disclosure. It is therefore contemplated that the appended claims shall cover any such modifications, or variations of the described embodiments as falling within the true spirit and scope of the invention.

10